




neodisher® Septo PreClean



Cleaning and disinfecting agent for the manual reprocessing of thermostable and thermolabile instruments



Liquid concentrate

Fields of application:

- Disinfecting pre-cleaning of thermostable and thermolabile instruments, dental instruments and flexible endoscopes e.g. prior to automated reprocessing for optimum personnel safety
- Wet disposal of surgical instruments and dental instruments over longer periods of time, e.g. overnight or over the weekend
- Also suitable for the manual cleaning and disinfection of thermostable and thermolabile instruments, including flexible endoscopes

Performance spectrum:

- Bactericidal and yeasticidal activity according to VAH¹ methods and European standards², also active against enveloped viruses according to European standards² and RKI/DVV³. Confirmed by certification
- Suitable for stainless steel, anodised aluminium, non-ferrous metals (copper and brass), silicone, plastics
- VAH¹-listed, On the IHO⁴-disinfectant list

Special properties:

- Excellent cleaning performance, non-protein-fixing
- Excellent material compatibility
- Free of aldehydes and quaternary ammonium compounds

Application and dosage:

neodisher Septo PreClean is used in immersion and ultrasonic baths. Depending on the desired activity the working solution must be prepared according to the dosing recommendations. Place the instruments resp. endoscopes without air bubbles into the working solution and adhere to the below-mentioned contact times.

For preparing the neodisher Septo PreClean working solution we recommend the use of softened water or soft water of less than 10 °d (1.8 mmol/l). The water hardness should not exceed 20 °d (3.6 mmol/l).

For wet disposal instruments can remain in the neodisher Septo PreClean working solution for up to 72 hours.

Disinfecting performance (20 °C, dirty conditions)	
Bactericidal (EN 13727, EN 14561, VAH)	5 ml/l (0.5 %), 15 min
Yeasticidal (EN 13624, EN 14562, VAH)	5 ml/l (0.5 %), 15 min
Active against enveloped viruses (incl. HIV, HBV, HCV) (EN 14476, EN 17111)	15 ml/l (1.5 %), 15 min
Active against enveloped viruses (incl. HIV, HBV, HCV) (RKI/DVV)	10 ml/l (1.0 %), 15 min

The unused working solution of neodisher Septo PreClean can be used for up to 7 days. Contaminated solutions however must be replaced daily or immediately if visibly dirty in accordance with the guideline of the Robert Koch Institute.

The neodisher Septo PreClean working solution has to be rinsed off completely with water (preferably deionised water). Small residues of neodisher Septo PreClean, which could remain on the instruments and be carried into the washer disinfector, neither cause foam nor a pressure loss in circulation pumps.

Aldehydic products should not be used before or after disinfection with neodisher Septo PreClean.

General instructions on use:

- For professional use only!
- For economical and controlled dosage, dosing ancillaries such as dosing device neomatik mediDOS are available. Please contact us.
- It is generally recommended to wear gloves when working with disinfectants.
- Reprocessing should comply with all ordinances pursuant to the regulations on medical devices and should be performed with appropriate validated processes.



neodisher[®] Septo PreClean

- Please observe the instrument manufacturer's recommendations for reprocessing according to the requirements of DIN EN ISO 17664.
- Not suitable for high-level disinfection of medical devices
- Do not mix with other products.

Expert reports:

The disinfecting activity has been confirmed by certification. Expert reports are available on request.

Technical data:

pH-range	8.9 - 8.7 (5 - 15 ml/l, determined in deionised water, 20 °C)
Viscosity	< 50 mPa s (concentrate, 20 °C)
Density	approx. 1.0 g/cm ³ (20 °C)

Ingredients:

Ingredients according to Regulation (EC) No 648/2004 on detergents:

< 5 % non-ionic surfactants
5 - 15 % amphoteric surfactants
also disinfectants, perfumes (limonene)


Active substances in 100 g:
8.0 g laurylpropylene diamine

CE-mark: 

neodisher Septo PreClean complies with European guidelines for medical devices.

If a serious incident occurs with the product, report it to the manufacturer and the relevant national authority.

Storage information:

Always store at a temperature between 0 °C and 30 °C. Usable for 3 years when stored as recommended. For expiry date refer to the stamp mark on the label behind the hourglass symbol .

Changes in the colour of the product may occur when storing in factory-sealed trade units. This has no impact on the properties of the product which are relevant for application.

Hazard and precautionary statements:

For safety information see Safety Data Sheets. These are available at www.drweigert.com under the headline "Service/Downloads".

Dispose only when container is empty and closed. For disposal of product residues, refer to Safety Data Sheet.

MB 4503/3-5
Revision Date: 07/2020

1 Verbund für Angewandte Hygiene [Association of Applied Hygiene], (VAH) Germany
2 according to European Standards EN13727, EN13624, EN14561, EN14562, EN 14476, EN 17111
3 according to the guidelines of the Robert Koch-Institute (RKI) and the Deutschen Vereinigung zur Bekämpfung von Viruskrankheiten (DVV) [German Association for the Control of Virus Diseases]
4 Industrieverband Hygiene und Oberflächenschutz [Industry Association Surface Protection and Hygiene], (IHO), Germany

With the above information, to our current knowledge we describe our product regarding safety necessities, but we do not involve any quality description or promise certain properties.